



**TANGO**  
Device  
Server

# **Novelec MCCE-2 Electrometers User's Guide**

## **MCCE2Electrometers Class**

**Revision: release\_2\_0\_2 - Author: xavela**  
**Implemented in C++**

### **Introduction:**

This class allows you to control all MCCE-2 Novelec electrometers through a Serial bus.

Supported types :

Novelec Electrometers : MCCE2

### **Class Inheritance:**

- Tango::Device\_3Impl
  - MCCE2Electrometers

## Properties:

Device Properties		
Property name	Property type	Description
<b>CommunicationLinkName</b>	Tango::DEV_STRING	The name of the device which manage the communication.
<b>ElectrometerType</b>	Tango::DEV_USHORT	The elctrometer type number. -> for Type 1 Novelec model : 1

### Device Properties Default Values:

Property Name	Default Values
CommunicationLinkName	No default value
ElectrometerType	No default value

**There is no Class properties.**

## Attributes:

Scalar Attributes			
Attribute name	Data Type	R/W Type	Expert
<b>electrometerChannel</b> : MCCE2 electrometer have two channels to support two amplifiers. WARN : default is 0 (please set this value first).	DEV_USHORT	WRITE	No

## Commands:

More Details on commands...

## Device Commands for Operator Level

Command name	Argument In	Argument Out
<b>Init</b>	DEV_VOID	DEV_VOID
<b>State</b>	DEV_VOID	DEV_STATE
<b>Status</b>	DEV_VOID	CONST_DEV_STRING
<b>RangeUP</b>	DEV_VOID	DEV_VOID
<b>RangeDOWN</b>	DEV_VOID	DEV_VOID
<b>SetZeroVFFunction</b>	DEV_VOID	DEV_VOID
<b>SetOffsetZeroV1Function</b>	DEV_VOID	DEV_VOID
<b>SetLeakageZeroV2Function</b>	DEV_VOID	DEV_VOID
<b>SetTestFunction</b>	DEV_VOID	DEV_VOID
<b>SetMeasureFunction</b>	DEV_VOID	DEV_VOID
<b>MCCE2_ON</b>	DEV_VOID	DEV_VOID
<b>MCCE2_OFF</b>	DEV_VOID	DEV_VOID
<b>Local</b>	DEV_VOID	DEV_VOID
<b>Remote</b>	DEV_VOID	DEV_VOID
<b>Reset</b>	DEV_VOID	DEV_VOID
<b>ClearRegisters</b>	DEV_VOID	DEV_VOID
<b>GetRange</b>	DEV_VOID	CONST_DEV_STRING
<b>SetRange</b>	DEV_STRING	DEV_VOID
<b>GetMode</b>	DEV_VOID	CONST_DEV_STRING
<b>GetPolarity</b>	DEV_VOID	CONST_DEV_STRING

### 1 - Init

- **Description:** This commands re-initialise a device keeping the same network connection.  
After an Init command executed on a device, it is not necessary for client to re-connect to the device.  
This command first calls the device *delete\_device()* method and then execute its *init\_device()* method.  
For C++ device server, all the memory allocated in the *nit\_device()* method must be freed in the *delete\_device()* method.  
The language device desctructor automatically calls the *delete\_device()* method.
- **Argin:**  
**DEV\_VOID** : none.
- **Argout:**  
**DEV\_VOID** : none.
- **Command allowed for:**

## 2 - State

- **Description:** This command gets the device state (stored in its *device\_state* data member) and returns it to the caller.
- **Argin:**  
**DEV\_VOID** : none.
- **Argout:**  
**DEV\_STATE** : State Code
- **Command allowed for:**

## 3 - Status

- **Description:** This command gets the device status (stored in its *device\_status* data member) and returns it to the caller.
- **Argin:**  
**DEV\_VOID** : none.
- **Argout:**  
**CONST\_DEV\_STRING** : Status description
- **Command allowed for:**

## 4 - RangeUP

- **Description:** Up the range of the electrometer. Throw : - electrometer::ElectrometerException if the range limit is reached  
- Tango::DevFailed if the command cannot be performed
- **Argin:**  
**DEV\_VOID** : no argin
- **Argout:**  
**DEV\_VOID** : no argout
- **Command allowed for:**

## 5 - RangeDOWN

- **Description:** Down the range of the electrometer. Throw : - electrometer::ElectrometerException if the range is negative -  
Tango::DevFailed if the command cannot be performed
- **Argin:**  
**DEV\_VOID** : no argin
- **Argout:**  
**DEV\_VOID** : no argout
- **Command allowed for:**

## 6 - SetZeroVFunction

- **Description:** Enable the Zero V/F MCCE2 mode.
- **Argin:**  
DEV\_VOID : no argin
- **Argout:**  
DEV\_VOID : no argout
- **Command allowed for:**

## 7 - SetOffsetZeroV1Function

- **Description:** Enable the Offset or Zero V1 MCCE2 mode.
- **Argin:**  
DEV\_VOID : no argin
- **Argout:**  
DEV\_VOID : no argout
- **Command allowed for:**

## 8 - SetLeakageZeroV2Function

- **Description:** Enable the Leakage or Zero V2 MCCE2 mode.
- **Argin:**  
DEV\_VOID : no argin
- **Argout:**  
DEV\_VOID : no argout
- **Command allowed for:**

## 9 - SetTestFunction

- **Description:** Enable theTest MCCE2 mode.
- **Argin:**  
DEV\_VOID : no argin
- **Argout:**  
DEV\_VOID : no argout
- **Command allowed for:**

## 10 - SetMeasureFunction

- **Description:** Enable the Measure MCCE2 mode.
- **Argin:**  
DEV\_VOID : no argin
- **Argout:**  
DEV\_VOID : no argout
- **Command allowed for:**

## 11 - MCCE2\_ON

- **Description:** Enable the MEASURE function
- **Argin:**  
DEV\_VOID : no argin
- **Argout:**  
DEV\_VOID : no argout
- **Command allowed for:**

## 12 - MCCE2\_OFF

- **Description:** Disable the MEASURE function
- **Argin:**  
DEV\_VOID : no argin
- **Argout:**  
DEV\_VOID : no argout
- **Command allowed for:**

## 13 - Local

- **Description:** Enable local keyboard and Key
- **Argin:**  
DEV\_VOID : no argin
- **Argout:**  
DEV\_VOID : no argout
- **Command allowed for:**

## 14 - Remote

- **Description:** Disable \"MODIFY\" and the \"ON-OFF\" keys
- **Argin:**  
DEV\_VOID : no argin
- **Argout:**  
DEV\_VOID : no argout
- **Command allowed for:**

## 15 - Reset

- **Description:** Restart the MCCE-2
- **Argin:**  
DEV\_VOID : no argin
- **Argout:**  
DEV\_VOID : no argout
- **Command allowed for:**

## 16 - ClearRegisters

- **Description:** Clear error(s)
- **Argin:**  
DEV\_VOID : no argin
- **Argout:**  
DEV\_VOID : no argout
- **Command allowed for:**

## 17 - GetRange

- **Description:** Returns the electrometer range.
- **Argin:**  
DEV\_VOID : no argin
- **Argout:**  
CONST\_DEV\_STRING : The actual electrometer range
- **Command allowed for:**

## 18 - SetRange

- **Description:** Apply the specified argin range, on the electrometer, if well formatted. Else an exception is thrown.
- **Argin:**  
**DEV\_STRING** : The range to apply on the electrometer
- **Argout:**  
**DEV\_VOID** : no argout
- **Command allowed for:**

## 19 - GetMode

- **Description:** Returns the electrometer mode which can be one of the following values : MEASURE, V/F Zero, OFFSET, LEAKAGE or TEST.
- **Argin:**  
**DEV\_VOID** : no argin
- **Argout:**  
**CONST\_DEV\_STRING** : The electrometer mode (MEASURE, LEAKAGE ...)
- **Command allowed for:**

## 20 - GetPolarity

- **Description:** Returns the electrometer polarity, which can be POSITIVE or NEGATIVE.
- **Argin:**  
**DEV\_VOID** : no argin
- **Argout:**  
**CONST\_DEV\_STRING** : The electrometer polarity
- **Command allowed for:**

---

**ESRF - Software Engineering Group**





**TANGO**  
Device  
Server

# **Novelec MCCE-2 Electrometers User's Guide**

## **MCCE2Electrometers Class**

**Revision: release\_2\_0\_2 - Author: xavela**  
**Implemented in C++**

### **Introduction:**

This class allows you to control all MCCE-2 Novelec electrometers through a Serial bus.

Supported types :

Novelec Electrometers : MCCE2

### **Class Inheritance:**

- Tango::Device\_3Impl
  - MCCE2Electrometers

## Properties:

Device Properties		
Property name	Property type	Description
<b>CommunicationLinkName</b>	Tango::DEV_STRING	The name of the device which manage the communication.
<b>ElectrometerType</b>	Tango::DEV_USHORT	The elctrometer type number. -> for Type 1 Novelec model : 1

## Device Properties Default Values:

Property Name	Default Values
CommunicationLinkName	No default value
ElectrometerType	No default value

**There is no Class properties.**

## Attributes:

Scalar Attributes			
Attribute name	Data Type	R/W Type	Expert
<b>electrometerChannel</b> : MCCE2 electrometer have two channels to support two amplifiers. WARN : default is 0 (please set this value first).	DEV_USHORT	WRITE	No

## Commands:

More Details on commands...